


Separation of Mianserin on Epitomize CSP-1A by UHPLC



Sample	Mianserin [D21800]		
System	Waters Acquity, H-Class		
Column	Epitomize CSP-1A, 3 micron, 3.0 x 50 mm		
Mobile Phase	80/20/0.1 Heptane/2-propanol/diethylamine		
Flow Rate	0.50 mL/min	Rt1 & Rt2 (min)	0.69 & 0.89
Temperature	25 °C	k1 & k2	0.45 & 0.87
Detection	UV at 254 nm	Separation Value	1.94
Injection	0.50 µL (1 mg/mL)	Resolution	2.98