


Separation of Oxprenolol on Epitomize CSP-1K by UHPLC



Sample	Oxprenolol [D26450]		
System	Waters Acquity, H-Class		
Column	Epitomize CSP-1K, 3 micron, 3.0 x 50 mm		
Mobile Phase	80/20/0.1 Heptane/2-propanol/DEA		
Flow Rate	0.50 mL/min	Rt1 & Rt2 (min)	0.89 & 1.51
Temperature	25 °C	k1 & k2	0.73 & 1.95
Detection	UV at 230 nm	Separation Value	2.68
Injection	0.50 µL (1 mg/mL)	Resolution	5.91