


## Separation of Oxprenolol on Epitomize CSP-1A by UHPLC



<b>Sample</b>	Oxprenolol	[D21961]
<b>System</b>	Waters Acquity, H-Class	
<b>Column</b>	Epitomize CSP-1A, 3 micron, 3.0 x 50 mm	
<b>Mobile Phase</b>	80/20/0.1 Heptane/2-propanol/diethylamine	
<b>Flow Rate</b>	0.50 mL/min	<b>Rt1 &amp; Rt2 (min)</b> 0.86 & 1.02
<b>Temperature</b>	25 °C	<b>k1 &amp; k2</b> 0.81 & 1.15
<b>Detection</b>	UV at 230 nm	<b>Separation Value</b> 1.42
<b>Injection</b>	2.0 µL (1 mg/mL)	<b>Resolution</b> 2.45