


Separation of 2-Phenoxypropionic Acid on Epitomize CSP-1A by UHPLC



Sample	2-Phenoxypropionic Acid [D20272]		
System	Waters Acquity, H-Class		
Column	Epitomize CSP-1A, 3 micron, 3.0 x 50 mm		
Mobile Phase	95/5/0.1 Heptane/2-propanol/TFA		
Flow Rate	0.50 mL/min	Rt1 & Rt2 (min)	1.56 & 2.07
Temperature	25 °C	k1 & k2	2.29 & 3.36
Detection	UV at 210 nm	Separation Value	1.47
Injection	1.50 µL (1 mg/mL)	Resolution	3.96