Chiral Applications Library

Application Note: UCEP-1Z-18061

Separation of Proglumide on Epitomize CSP-1Z by UHPLC



Sample	Proglumide [D20723]		
System	Waters Acquity, H-Class		
Column	Epitomize CSP-1Z, 3 micron, 3.0 x 50 mm		
Mobile Phase	90/10/0.1 Heptane/2-propanol/TFA		
Flow Rate	0.50 mL/min	**Rt1 & Rt2 (min)**	1.70 & 3.35
Temperature	25 °C	**k1 & k2**	2.30 & 5.48
Detection	UV at 230 nm	**Separation Value**	2.39
Injection	0.50 µL (1 mg/mL)	**Resolution**	5.18

Orochem Technologies Inc
340 Shuman Blvd.
Naperville, IL 60563

Tel.: 630-210-8300
Fax: 630-210-83150250
www.orochem.com