


Separation of 1,2,3,4-Tetrahydro-1-naphthol on Epitomize CSP-1C by UHPLC



Sample	1,2,3,4-Tetrahydro-1-naphthol [D18712]		
System	Waters Acquity, H-Class		
Column	Epitomize CSP-1C, 3 micron, 3.0 x 50 mm		
Mobile Phase	97/3 Heptane/2-propanol		
Flow Rate	0.50 mL/min	Rt1 & Rt2 (min)	2.53 & 2.77
Temperature	25 °C	k1 & k2	3.96 & 4.42
Detection	UV at 230 nm	Separation Value	1.12
Injection	5.00 µL (1 mg/mL)	Resolution	1.43