


Separation of 1,2,3,4-Tetrahydro-1-naphthol on Epitomize CSP-1J by UHPLC



Sample	1,2,3,4-Tetrahydro-1-naphthol [D19234]		
System	Waters Acquity, H-Class		
Column	Epitomize CSP-1J, 3 micron, 3.0 x 50 mm		
Mobile Phase	95/5 Heptane/2-propanol		
Flow Rate	0.50 mL/min	Rt1 & Rt2 (min)	1.43 & 1.58
Temperature	25 °C	k1 & k2	1.80 & 2.10
Detection	UV at 254 nm	Separation Value	1.17
Injection	5.00 µL (1 mg/mL)	Resolution	1.53